

OQ Protocol



Company: Moscow State University of
Operator: Default
Instrument Type: Senterra Microscope 785 nm Grating 1200
Optics Configuration: Microscope with: 785 nm, Grating 1200, CCD
Accessory: None
Instrument Serial Number: Result
Instrument Firmware Version:
OPUS/DB Version: 20090227 / 6.5.9
Overall Test Result: FAIL
Test expires: Test failed
Test Date/Time: 2012/01/18 18:19:58
Test Spectra Path: C:\Program Files\OPUS_65\Validation\Data\20120118\181958
Comment:

Throughput Test

Throughput minimum requirement: 3700 Throughput measured: 3199

Overall Test Result = FAIL



Date and Signature

Date and Signature